

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY Goodyear  
DATE 3/1/2024 TIME 6:50  
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample - Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE K Miyazawa 1139

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-003795

Test Date: 03/01/2024

Test Time: 05:22

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 011

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	05:22:46
Calibration Chk	0.098	05:23:07
Air Blank	0.000	05:23:47

No RFI Detected

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37950301240687

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 01/31/2024  
Last Annual Maintenance: 08/07/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 03/01/2024  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 04:45  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 24923100A1  
Expiration Date: 10/05/2025  
Bottle #: 011  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

## RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect


## OPERATOR COMMENTS

All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .

Test	g/210L	Time
Air Blank	0.000	05:08:36
Diagnostic Test	PASS	05:09:12
Air Blank	0.000	05:09:49
Calibration Chk	0.099	05:10:11
Air Blank	0.000	05:10:50
Subject Test 1	DEF*	05:14:03
Air Blank	0.000	05:14:47
Air Blank	0.000	05:15:29
Subject Test 2	0.000	05:15:47
Air Blank	0.000	05:16:27
Wait		05:20:31
Air Blank	0.000	05:21:08
Subject Test 3	INV**	05:21:35
Air Blank	RFI***	05:21:54

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

Not a Successfully Completed Test Sequence



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

## Certificate of Analysis

Certificate ID: 15958  
Part #: BAC105L100T  
Cylinder Size: 105L  
Lot Number: 24923100A1  
Expiration: 10/5/2025

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	260 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:

CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

\*Traceable to:  
Certified Reference Material - 261.0 µmol/mol  
Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

09-26-2023  
Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-003795

Test Date: 03/07/2024

Test Time: 03:21

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 056

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	03:22:14
Calibration Chk	0.099	03:22:35
Air Blank	0.000	03:23:15

No RFI Detected

Successfully Completed Test Sequence

**From:** Erin Boone  
**Sent:** Tuesday, July 30, 2024 3:28 PM  
**To:** CityProsecutor  
**Subject:** RE: Intoxilyzer 9000

**⚠ This email arrived from an external source - Please exercise caution when opening any attachments or clicking on links**

90-004605 was received into the lab on June 4, 2024. We are waiting on parts to complete this instrument before it can be released.

90-003795 was received into the lab on March 20, 2024 and sent to CMI on July 12, 2024 for repairs.

If you need anything further, let me know.

Thanks,

Erin E. Boone  
Forensic Scientist IV  
AZ Dept. of Public Safety  
Central Regional Crime Lab



ARIZONA DEPARTMENT OF PUBLIC SAFETY  
 CRIMINAL JUSTICE SUPPORT DIVISION  
 INTOXILYZER SERVICE REQUEST

Admin Use Only
CR#
Date: 3/20/24
Initials: [Signature]

<input checked="" type="checkbox"/> Repair	Date Submitted: 3/20/2024	Intoxilyzer Serial Number: 90-003795
<input type="checkbox"/> Preventative Maintenance	Submitting Agency Name: GOODYEAR	Location of Equipment/Agency: Goodyear P.D.
Last Name: MIYAZATO	First Name: Kory	Badge: 1139
Cell or Primary Phone: 623-692-2492	Work Phone: 623-882-7468	Email: KMIYAZATO@GOODYEAR.AZ.GOV

Other Equipment Sent:

- Gas Cylinder
- Power Cord

Problem Description (if applicable) (please be specific)

IRPCM STATUS ERROR