

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear  
DATE 3/1/2024 TIME 0542  
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.102 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGNATURE K Miyazato 1139

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-004605

Test Date: 03/01/2024

Test Time: 05:56

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #:

Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Operator: MIYAZATO, KORY

Permit #:30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	05:57:12
Calibration Chk	0.102	05:57:34
Air Blank	0.000	05:58:13

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL  
ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000  
Serial Number: 90-004605  
Software Version: 9439.01.00  
Analytical Report Number: 46050301240083  
QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 01/31/2024  
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Test Date: 03/01/2024  
Name: 31-Day Check,  
Driver's License #:  
Weight:  
Gender:  
State of Issue:  
Date of Birth:  
Driver's License Expiration:  
Age:  
Deprivation Start Time: 05:00  
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD  
Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 24923100A1  
Expiration Date: 10/05/2025  
Bottle #:  
Last Changed By:  
GUILBAULT, MICHAEL  
Permit #: 30305

Test	g/210L	Time
Air Blank	0.000	05:42:59
Diagnostic Test	PASS	05:43:35
Air Blank	0.000	05:44:12
Calibration Chk	0.101	05:44:33
Air Blank	0.000	05:45:13
Subject Test 1	DEF*	05:48:35
Air Blank	0.000	05:49:21
Air Blank	0.000	05:50:01
Subject Test 2	0.000	05:50:19
Air Blank	0.000	05:50:59
Wait		05:55:04
Air Blank	0.000	05:55:41
Subject Test 3	INV**	05:56:02
Air Blank	RFI***	05:56:22

\*Deficient Sample

\*\*Invalid Sample - Begin new deprivation period

\*\*\*RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

\* Deficient Sample

\*\* Invalid Sample - Begin new deprivation period

\*\*\* RFI Detect

All air blank results must be 0.000.

Consecutive subject test results must not differ  
by more than 0.020 g/210L.

Standard check results must be  $\pm 10\%$ .

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear  
DATE 3/31/24 TIME 2304  
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check  
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
5. Standard Calibration Check:  
0.101 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGNATURE K Miyazato 1139

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-004605  
Software Version: 9439.01.00  
Analytical Report Number: 46050331240110

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 03/01/2024  
Last Annual Maintenance: 08/31/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 03/31/2024  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 22:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 24923100A1  
Expiration Date: 10/05/2025  
Bottle #:  
Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Test	g/210L	Time
Air Blank	0.000	23:04:46
Diagnostic Test	PASS	23:05:22
Air Blank	0.000	23:05:59
Calibration Chk	0.101	23:06:20
Air Blank	0.000	23:07:00
Subject Test 1	DEF*	23:10:12
Air Blank	0.000	23:10:54
Air Blank	0.000	23:11:34
Subject Test 2	0.000	23:11:52
Air Blank	0.000	23:12:34
Wait		23:16:37
Air Blank	0.000	23:17:14
Subject Test 3	INV**	23:17:53
Air Blank	RFI***	23:18:03

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 03/31/2024

Test Time: 23:18

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #:

Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	23:18:56
Calibration Chk	0.101	23:19:18
Air Blank	0.000	23:19:58

No RFI Detected