

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa AGENCY Goodyear
DATE 5/1/24 TIME 2129
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. _____ AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ___ NO ___

COMMENTS:

SIGNATURE K Miyazawa 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050501240150

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 03/31/2024
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 05/01/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #:
Last Changed By: GUILBAULT, MICHAEL

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:29:36
Diagnostic Test	PASS	21:30:12
Air Blank	0.000	21:30:49
Calibration Chk	0.102	21:31:11
Air Blank	0.000	21:31:50
Subject Test 1	DEF*	21:35:02
Air Blank	0.000	21:35:52
Air Blank	0.000	21:36:34
Subject Test 2	0.000	21:36:51
Air Blank	0.000	21:37:30
Wait		21:41:34
Air Blank	0.000	21:42:11
Subject Test 3	INV**	21:42:32
Air Blank	RFI***	21:42:48

Permit #: 30305

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 05/01/2024

Test Time: 21:43

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #:

Last Changed By: GUILBAULT, MICHAEL

Permit #: 30305

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:43:39
Calibration Chk	0.101	21:44:01
Air Blank	0.000	21:44:41

No RFI Detected

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 05/29/2024

Test Time: 20:21

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 51

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	20:22:23
Calibration Chk	0.102	20:22:45
Air Blank	0.000	20:23:24

No RFI Detected

Successfully Completed Test Sequence



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 15958
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 24923100A1
Expiration: 10/5/2025

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:	Distributed by:
Ethanol	260 ppm	+/-0.002 BAC(G/210L)	NDIR	CMI Inc. 316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com
Nitrogen	Balance	[5.2 ppm]		

*Traceable to:
Certified Reference Material - 261.0 µmol/mol
Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).


Specialty Gas Lab Tech

09-26-2023
Issuance Date


J. Lanorio



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050531240198

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 05/01/2024
Last Annual Maintenance: 08/31/2023

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 05/31/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 51
Last Changed By: MIYAZATO, KORY

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	22:05:39
Diagnostic Test	DIA*	22:05:44
Air Blank	0.000	22:06:23

*00004000000
IRPCM Status Error

Permit #: 30215

RESULTS

EXCEPTION MESSAGES

* Diagnostic Fail

OPERATOR COMMENTS

No RFI Detected

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 05/31/2024

Test Time: 22:30

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 51

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	22:31:12
Reference	UNS*	22:31:18
Air Blank	0.000	22:31:55

*Unstable Signal

No RFI Detected