

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/04/2024

19:37:53

-----  
Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139  
-----

QAS: K. MIYAZATO #1139

GOODYEAR PD  
-----

Operator: K. MIYAZATO #1139

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	19:38:24
0.100 Cal Check	0.100	19:38:45
Air Blank	0.000	19:39:15

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K. MAZATTO AGENCY GOODYEAR  
DATE 6/4/24 TIME 1940  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145 TH AVE

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.  
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.  
() 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
() 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

SIGNATURE [Signature] 1139

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
06/04/2024

19:40:45

---

Standard Lot#: 24923100A1  
Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139  
GOODYEAR PD

---

Operator: K. MIYAZATO #1139  
GOODYEAR PD

---

Subject: SQAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	19:41:32
Diagnostic Test	Pass	19:42:07
Air Blank	0.000	19:42:35
0.100 Cal Check	0.100	19:42:55
Air Blank	0.000	19:43:26
Subject Test	0.000	19:43:37
Air Blank	ABT*	19:43:44
Air Blank	0.000	19:44:12

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/04/2024

19:44:14

---

Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	19:44:47
Air Blank	0.000	19:45:16

\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: 11 N 145TH AVE  
Serial Number: 80-000312  
Core Version: 8105.48  
06/04/2024

19:45:19

---

Standard Lot#: 24923100A1  
Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139  
GOODYEAR PD

---

Operator: K. MIYAZATO #1139  
GOODYEAR PD

---

Subject: SQAP,TEST  
DOB: 04/02/1973  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	19:45:46
Air Blank	0.000	19:46:15

\*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/04/2024

19:46:17

---

Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M

Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	19:47:05
Diagnostic Test	Pass	19:47:40
Air Blank	0.000	19:48:08
0.100 Cal Check	0.100	19:48:29
Air Blank	0.000	19:48:59
Subject Test	0.077*	19:52:06
Air Blank	0.000	19:52:38
Air Blank	0.000	19:53:06
Subject Test	XXX**	19:53:20
Air Blank	0.000	19:53:53

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/04/2024

19:54:04

-----  
Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139  
-----

QAS: K. MIYAZATO #1139

GOODYEAR PD  
-----

Operator: K. MIYAZATO #1139

GOODYEAR PD  
-----

Test	g/210L	Time
Air Blank	0.000	19:54:34
0.100 Cal Check	0.099	19:54:55
Air Blank	0.000	19:55:25

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K Miyazato AGENCY GOODYEAR  
DATE 6/4/24 TIME 1954  
INTOXILYZER SERIAL # 80-000312 LOCATION HN 145 TH AVO

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K Miyazato 1139



EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY Goodyear  
DATE 6/30/24 TIME 2007  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145 TH AVE

(  ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.  
(  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.  
(  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(  ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE

K Miyazawa 1139

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/30/2024

20:07:28

---

Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M

Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:08:55
Diagnostic Test	Pass	20:09:30
Air Blank	0.000	20:09:58
0.100 Cal Check	0.099	20:10:18
Air Blank	0.000	20:10:49
Subject Test	0.000	20:11:04
Air Blank	ABT*	20:11:10
Air Blank	0.000	20:11:39

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/30/2024

20:11:41

---

Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR

---

Subject: SQAP, TEST

DOB: 04/02/1973

Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	20:12:25
Air Blank	0.000	20:12:53

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/30/2024

20:14:03

---

Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M

Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	20:14:38
Air Blank	0.000	20:15:07

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/30/2024

20:15:11

---

Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

---

QAS: K. MIYAZATO #1139

GOODYEAR PD

---

Operator: K. MIYAZATO #1139

GOODYEAR PD

---

Subject: SQAP,TEST

DOB: 04/02/1973

Sex: M

Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:15:58
Diagnostic Test	Pass	20:16:33
Air Blank	0.000	20:17:00
0.100 Cal Check	0.099	20:17:21
Air Blank	0.000	20:17:51
Subject Test	0.014*	20:20:58
Air Blank	0.000	20:21:30
Air Blank	0.000	20:21:58
Subject Test	XXX**	20:22:13
Air Blank	0.000	20:22:46

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear  
DATE 6/30/24 TIME \_\_\_\_\_  
INTOXILYZER SERIAL # 80-000312 LOCATION 11 N 145 TH AVE

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K Miyazato 1139

INTOXILYZER 8000

Location: 11 N 145TH AVE

Serial Number: 80-000312

Core Version: 8105.48

06/30/2024

20:23:53

-----  
Standard Lot#: 24923100A1

Last Changed By: K. MIYAZATO #1139

-----  
QAS: K. MIYAZATO #1139

GOODYEAR PD

-----  
Operator: K. MIYAZATO #1139

GOODYEAR PD

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      20:24:24  
0.100 Cal Check                0.098                      20:24:44  
Air Blank                      0.000                      20:25:15

Preventative Maintenance  
Performed -- Timer Reset