

**From:** Erin Boone  
**Sent:** Tuesday, July 30, 2024 3:28 PM  
**To:** CityProsecutor  
**Subject:** RE: Intoxilyzer 9000

**⚠ This email arrived from an external source - Please exercise caution when opening any attachments or clicking on links**

90-004605 was received into the lab on June 4, 2024. We are waiting on parts to complete this instrument before it can be released.

90-003795 was received into the lab on March 20, 2024 and sent to CMI on July 12, 2024 for repairs.

If you need anything further, let me know.

Thanks,

Erin E. Boone  
Forensic Scientist IV  
AZ Dept. of Public Safety  
Central Regional Crime Lab

W



**ARIZONA DEPARTMENT OF PUBLIC SAFETY  
CRIMINAL JUSTICE SUPPORT DIVISION  
INTOXILYZER SERVICE REQUEST**

I90004605-0002  
CR# KDS 6/4/24

<input checked="" type="checkbox"/> Repair <input type="checkbox"/> Preventative Maintenance		Date Submitted 6/4/24	Intoxilyzer Serial Number I90004605
Submitting Agency Name GOODYEAR PD		Location of Equipment/Agency	
Last Name MIYAZATO		First Name KORY	Badge 1139
Cell or Primary Phone 623-692-2492	Work Phone 623-932-1220	Email KMIYAZATO@GOODYEARAZ.GOV	
Other Equipment Sent: <input type="checkbox"/> Gas Cylinder <input type="checkbox"/> Power Cord		Printer Used With 9000: <input type="checkbox"/> Internal <input checked="" type="checkbox"/> External	
Problem Description (if applicable) (please be specific)  IRPCM STATUS ERROR / DIAGNOSTIC FAIL  UNSTABLE SIGNAL RESPONSE			