

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear
DATE 9/30/24 TIME 2047
INTOXILYZER SERIAL # 90-004605

1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0.00 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0.101 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazato 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050930240200

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 05/31/2024
Last Annual Maintenance: 09/20/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 09/30/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 20:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 52
Last Changed By: MIYAZATO, KORY

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	20:47:08
Diagnostic Test	PASS	20:47:44
Air Blank	0.000	20:48:21
Calibration Chk	0.101	20:48:42
Air Blank	0.000	20:49:22
Subject Test 1	DEF*	20:52:35
Air Blank	0.000	20:53:17
Air Blank	0.000	20:53:57
Subject Test 2	0.000	20:54:15
Air Blank	0.000	20:54:56
Wait		20:59:00
Air Blank	0.000	20:59:37
Subject Test 3	INV**	21:00:13
Air Blank	RFI***	21:00:31

Permit #: 30215

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 09/30/2024

Test Time: 21:01

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 52

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:01:24
Calibration Chk	0.101	21:01:46
Air Blank	0.000	21:02:26

No RFI Detected

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 09/30/2024

Test Time: 20:54

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 52

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	20:54:49
Calibration Chk	0.103	20:55:10
Air Blank	0.000	20:55:50

No RFI Detected

Successfully Completed Test Sequence