

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY GOODYEAR
DATE 09/30/24 TIME 2:05
INTOXILYZER SERIAL # 90-005394

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.100 AC
 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
 5. Standard Calibration Check:
0.100 AC
 6. Air Blanks Completed
 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE K Miyazawa 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005394
Software Version: 9439.01.00
Analytical Report Number: 53940930240001

QAS:
QAS Permit #:
Agency:
Last 31-Day Check:
Last Annual Maintenance: 08/30/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 09/30/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 20:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 51
Last Changed By: MIYAZATO, KORY

Permit #: 30215

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	20:55:56
Diagnostic Test	PASS	20:56:32
Air Blank	0.000	20:57:09
Calibration Chk	0.099	20:57:30
Air Blank	0.000	20:58:10
Subject Test 1	DEF*	21:01:23
Air Blank	0.000	21:03:01
Air Blank	0.000	21:03:41
Subject Test 2	0.000	21:04:09
Air Blank	0.000	21:04:52
Wait		21:08:55
Air Blank	0.000	21:09:32
Subject Test 3	INV**	21:09:49
Air Blank	RFI***	21:10:06

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-005394

Test Date: 09/30/2024

Test Time: 21:10

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 51

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:10:58
Calibration Chk	0.100	21:11:20
Air Blank	0.000	21:12:00

No RFI Detected

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-005394

Test Date: 09/30/2024

Test Time: 20:51

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 51

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	20:51:34
Calibration Chk	0.100	20:51:55
Air Blank	0.000	20:52:35

No RFI Detected

Successfully Completed Test Sequence