



ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER MAINTENANCE REPORT

CR NO. I900037950006

Page 1 of 1

AGENCY Goodyear Police Department
 Goodyear, AZ 85338
OFFICER MIYAZATO, #1139
DATE October 17, 2024



INSTRUMENT

90003795

INSTRUMENT LOCATION

Goodyear Police Department

MAINTENANCE REQUESTED

Preventative Maintenance

MAINTENANCE PERFORMED

Preventative Maintenance Performed:

Calibration, Verification

Repaired at CMI.

A handwritten signature in black ink, appearing to read "Aleksis Croci".

ALEKSIS CROCI, #11246, Forensic Scientist
Central Regional Crime Laboratory
2102 W. Encanto Blvd., Phoenix, AZ 85009
(602) 223-2394 acroci@azdps.gov

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVENUE

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37951112240715
QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 03/01/2024
Last Annual Maintenance: 10/02/2024

SUBJECT INFORMATION

Test Date: 11/12/2024
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 23:00
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD
Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 2/20/2026
Bottle #: 19
Last Changed By: MIYAZATO, KORY
Permit #: 30215

Table with 3 columns: Test, g/210L, Time. Rows include Air Blank, Diagnostic Test, Calibration Chk, Subject Test 1, Subject Test 2, Subject Test 3, and various Air Blank tests.

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795
Test Date: 11/13/2024

Test Time: 23:07

Location: 11 N 145TH AVENUE

Software Version: 9439.01.00

Standard Lot #: 302-402972207
Standard Value: 0.100
Expiration Date: 02/20/2026
Bottle #: 19
Last Changed By: MIYAZATO, KORY
Permit #: 30215

Operator: MIYAZATO, KORY
Permit #: 30215
Agency: GOODYEAR PD

Table with 3 columns: Test, g/210L, Time. Rows include Air Blank, Calibration Chk, and Air Blank.

No RFI Detected

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazaki 1139 AGENCY GOODYEAR P.D.
DATE 11/12/24 TIME 2334
INTOXILYZER SERIAL # 90-003795

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. .100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE [Signature] 1139

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

CMI
316 EAST NINTH STREET
OWENSBORO, KY 42303

INVOICE # : 1127780136
PO # : 8005136
CMI ITEM # : 340287
Date: February 20, 2024

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
PRODUCT EXPIRATION: February 20, 2026

CALGAZ LOT#: 302-402972207
ETHANOL IN NITROGEN

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

*Certification traceability is recognized by NIST (USA) through the CIPM (International Committee of Weights and Measures) MRA (Mutual Recognition Arrangement)

CALGAZ CYLINDER SIZE: 6DM

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251